

JADUAL HARGA (SEBUT HARGA)

<p>5. AIRFLOW VOLUME :</p> <p>MINIMUM :</p> <p>a) INFLOW : 346 cmh (204 cfm)</p> <p>b) DOWNFLOW : 764 cmh (449 cfm)</p> <p>c) EXHAUST : 346 cmh (204 cfm)</p> <p>6. <u>FILTER EFFICIENCY (%) :</u></p> <p>MINIMUM ULPA/HEPA FILTRATION SYSTEM OPERATE AT A TYPICAL EFFICIENCY >99.995% AT 0.1 TO 0.3 MICRON SIZES.</p> <p>7. <u>SOUND EMISSION (IEST-RP-CC002.2)*</u></p> <p>*NOISE READING IN OPEN FIELD CONDITION / ANECHOIC CHAMBER MINIMUM ; NSF: 60 dBA / EN: 57 dBA</p> <p>8. <u>FLUORESCENT LAMP INTENSITY :</u></p> <p>MINIMUM \geq 900LUX</p> <p>9. <u>MICROPROCESSOR CONTROLLER :</u></p> <ul style="list-style-type: none"> - DIGITAL DISPLAY - ALARM FUNCTION AUDIBLE & VISUAL <p>10. <u>ELECTRICAL</u></p> <p>a) MAXIMUM POWER / CURRENT CONSUMPTION : 360 W / 2 A</p> <p>b) NOMINAL POWER CONSUMPTION : 308 W</p> <p>11. <u>CABINET CONSTRUCTION:</u></p> <p>a) <u>MAIN BODY :</u> EG STEEL WITH ISOCIDE OVEN-BAKED EPOXY-POLYESTER POWDER COATING</p> <p>b) <u>WORK ZONE :</u> STAINLESS STEEL, TYPE 304.</p>							
---	--	--	--	--	--	--	--

JADUAL HARGA (SEBUT HARGA)

<p>c) <u>SIDE WALL</u> : TEMPERED GLASS</p> <p>12. FEATURES:</p> <p><u>MINIMUM :</u></p> <p>a) FIXED STAND WITH CASTER WHEELS / LEVELING FEET - 28" /34".</p> <p>13. WARRANTY :</p> <p>MINIMUM 1 YEARS WARRANTY AGAINST MANUFACTURER DEFECTS</p> <p>14. ON SITE TRAINING :</p> <p>WITH CERTIFICATION FOR THE PARTICIPANTS AFTER TRAINING</p> <p>15. YEARLY RE-CERTIFICATE FOR BIOSAFETY CABINET :</p> <p>MINIMUM 1 YEARS</p>						<p>JUMLAH</p> <hr/> <hr/> <p>SST (%)</p> <p>JUMLAH BESAR</p> <hr/> <hr/>		
--	--	--	--	--	--	--	--	--

JADUAL HARGA (SEBUT HARGA)

Jika berkaitan, sila nyatakan:

1. Kos Penyelenggaraan termasuk alat ganti selepas tempoh jaminan:

Tahun 1: RM _____

Tahun 2: RM _____

Tahun 3: RM _____

Tahun 4: RM _____

Tahun 5: RM _____

TANDATANGAN

NAMA

COP SYARIKAT

2. Tempoh Jaminan (minima 1 Tahun): _____

3. Wajib diisi bagi bekalan ex-stock:

a) Tarikh bekalan tiba di Malaysia atau mula disimpan oleh syarikat _____

b) Lampirkan salinan borang Kastam